

Amplite™ Colorimetric Xanthine Assay Kit

 Catalog number: 13842
 Unit size: 200 Tests

Component	Storage	Amount
Component A: Amplite™ Red Substrate (light sensitive)	Freeze (< -15 °C), Minimize light exposure	1 vial
Component B: Assay Buffer	Freeze (< -15 °C)	1 bottle (20 mL)
Component C: Horseradish Peroxidase (lyophilized)	Freeze (< -15 °C), Minimize light exposure	1 vial
Component D: Xanthine Standard	Freeze (< -15 °C), Minimize light exposure	1 vial (100 µL, 20 mM)
Component E: Xanthine Oxidase (lyophilized)	Freeze (< -15 °C), Minimize light exposure	1 vial
Component F: DMSO	Freeze (< -15 °C)	1 vial (100 µL)

OVERVIEW

Xanthine is a purine base found in most human body tissues and fluids. A number of stimulants are derived from xanthine, including caffeine, aminophylline, IBMX, paraxanthine, pentoxifylline, theobromine, and theophylline, which can stimulate heart rate, force of contraction, cardiac arrhythmias at high concentrations. Therefore, detection of Xanthine alteration in biological samples is important for disease diagnosis and therapy monitoring. Amplite™ Colorimetric Xanthine Assay Kit provides a quick and ultrasensitive method for the measurement of xanthine. It can be performed in a convenient 96-well or 384-well microtiter-plate format. Xanthine is oxidized to uric acid in the presence of xanthine oxidase to release hydrogen peroxide, which can be specifically measured with Amplite™ Red by an absorbance microplate reader at 576 nm. With Amplite™ Colorimetric Xanthine Assay Kit, as low as 1.2 µM xanthine was detected in a 100 µL reaction volume.

AT A GLANCE

Protocol Summary

1. Prepare xanthine standards and test samples (50 µL)
2. Add xanthine working solution (50 µL)
3. Incubate at room temperature for 30 - 60 min
4. Read absorbance increase at OD ratio of 570/610 nm

Important Thaw all the kit components at room temperature before starting the experiment.

KEY PARAMETERS

Absorbance microplate reader

Absorbance 570/610 nm
 Recommended plate Clear bottom

PREPARATION OF STOCK SOLUTIONS

Unless otherwise noted, all unused stock solutions should be divided into single-use aliquots and stored at -20 °C after preparation. Avoid repeated freeze-thaw cycles.

1. Amplite™ Red Substrate stock solution (250X)

Add 40 µL of DMSO (Component F) into the vial of Amplite™ Red Substrate (Component A).

Note Amplite™ Red Substrate is unstable in the presence of thiols such as dithiothreitol (DTT) and 2-mercaptoethanol. The final concentration of DTT or 2-mercaptoethanol in the reaction should be no higher than 10 µM. The assay should be performed at pH 7 – 8 (pH 7.4 is recommended) as Amplite™ Red is unstable pH >8.5.

2. HRP stock solution (500X)

Add 100 µL of Assay Buffer (Component B) into the vial of Horseradish Peroxidase (Component C).

3. Xanthine Oxidase (XO) stock solution (100X)

Add 100 µL of assay buffer (Component B) into the vial of Xanthine Oxidase Standard (Component E) to make Xanthine Oxidase (XO) stock solution (100X).

4. Xanthine standard solution (100 µM)

Add 5 µL of Xanthine Standard (Component D) into 995 µL of Assay Buffer (Component B) to get 100 µM Xanthine standard solution.

PREPARATION OF STANDARD SOLUTION

For convenience, use the Serial Dilution Planner:
<https://www.aatbio.com/tools/serial-dilution/13842>

Xanthine standard

Perform 1:3 serial dilution with 100 µM Xanthine standard solution (X7) to obtain Xanthine standards (X6 - X1).

PREPARATION OF WORKING SOLUTION

Add 20 µL of Amplite™ Red Substrate stock solution (250X), 10 µL of HRP stock solution (500X), and 50 µL of Xanthine Oxidase stock solution (100X) into 5 mL of Assay Buffer to make a total volume of 5.08 mL. Protect from light.

SAMPLE EXPERIMENTAL PROTOCOL

Table 1. Layout of Xanthine standards and test samples in a clear bottom 96-well microplate (50 µL/well). X= xanthine standards (X1 - X7, 0.1 to 100 µM), BL=blank control, TS=test samples.

BL	BL	TS	TS
X1	X1
X2	X2
X3	X3		
X4	X4		
X5	X5		
X6	X6		
X7	X7		

Table 2. Reagent composition for each well.


Well	Volume	Reagent
X1 - X7	50 µL	Serial Dilutions (0.1 to 100 µM)
BL	50 µL	Assay Buffer (Component B)
TS	50 µL	test sample

1. Prepare xanthine standards (X), blank controls (BL), and test samples (TS) according to the layout provided in Tables 1 and 2. For a 384-well plate, use 25 µL of reagent per well instead of 50 µL.
2. Add 50 µL of working solution to each well of xanthine standard, blank control, and test samples to make the total xanthine assay volume of 100 µL/well. For a 384-well plate, add 25 µL of working solution into each well instead, for a total volume of 50 µL/well.
3. Incubate the reaction for 30 to 60 minutes at room temperature, protected from light.
4. Monitor signal intensity with an absorbance plate reader at OD ratio of 570/610 nm.

EXAMPLE DATA ANALYSIS AND FIGURES

The reading (Abs 570/Abs 610) obtained from the blank standard well is used as a negative control. Subtract this value from the other standards' readings to obtain the base-line corrected values. Then, plot the standards' readings to obtain a standard curve and equation. This equation can be used to calculate Xanthine samples. We recommend using the Online Linear Regression Calculator which can be found at:

<https://www.aatbio.com/tools/linear-logarithmic-semi-log-regression-online-calculator>



Xanthine dose response was measured with Amplitude™ Colorimetric Xanthine Assay Kit in a 96-well clear bottom plate using a SpectraMax reader (Molecular Devices).

Image generated with Quest™ Graph, ©2018 AAT Bioquest

Figure 1. Xanthine dose response was measured with Amplitude™ Colorimetric Xanthine Assay Kit in a 96-well clear bottom plate using a SpectraMax reader (Molecular Devices).

DISCLAIMER

AAT Bioquest provides high-quality reagents and materials for research use only. For proper handling of potentially hazardous chemicals, please consult the Safety Data Sheet (SDS) provided for the product. Chemical analysis and/or reverse engineering of any kit or its components is strictly prohibited without written permission from AAT Bioquest. Please call 408-733-1055 or email info@aatbio.com if you have any questions.