

Indo-1, AM *UltraPure Grade* *CAS 112926-02-0*

 Catalog number: 21032, 21036
 Unit size: 1 mg, 20x50 ug

Component	Storage	Amount (Cat No. 21032)	Amount (Cat No. 21036)
Indo-1, AM *UltraPure Grade* *CAS 112926-02-0*	Freeze (< -15 °C), Minimize light exposure	1 vial (1 mg)	20x50 ug

OVERVIEW

Calcium measurements are critical for numerous biological investigations. Fluorescent probes that show spectral responses upon binding to Ca²⁺ have enabled researchers to investigate changes in intracellular free Ca²⁺ concentrations by using fluorescence microscopy, flow cytometry, fluorescence spectroscopy and fluorescence microplate readers. This cell-permeant Indo-1 AM *UltraPure Grade*, is a UV light excitable, emission ratioable Ca²⁺ indicator. Upon binding to Ca²⁺, the emission maximum of Indo-1 AM shifts from 480 nm to 400 nm. Indo-1 is preferred for flow cytometry, in which it is more practical to use a single laser for excitation, such as the 351-364 nm spectral lines of the argon-ion laser.

KEY PARAMETERS
Fluorescence microscope

Excitation	Indo-1 filter set
Emission	Indo-1 filter set
Recommended plate	Black wall/clear bottom

Fluorescence microplate reader

Excitation	340
Emission	400, 475
Cutoff	Ex/Em = 340/400, no cut off. Ex/Em = 340/475, cut off 455
Recommended plate	Black wall/clear bottom
Instrument specification(s)	Bottom read mode/Programmable liquid handling

PREPARATION OF STOCK SOLUTIONS

Unless otherwise noted, all unused stock solutions should be divided into single-use aliquots and stored at -20 °C after preparation. Avoid repeated freeze-thaw cycles.

Indo-1 AM *UltraPure Grade* Stock Solution

Prepare a 2 to 5 mM stock solution of Indo-1 AM in high-quality, anhydrous DMSO.

PREPARATION OF WORKING SOLUTION
Indo-1 AM *UltraPure Grade* Working Solution

On the day of the experiment, either dissolve Indo-1 AM in DMSO or thaw an aliquot of the indicator stock solution to room temperature. Prepare a dye working solution of 2 to 20 μM in a buffer of your choice (e.g., Hanks and Hepes buffer) with 0.04% Pluronic® F-127. For most cell lines, Indo-1 AM at a final concentration of 4-5 μM is recommended. The exact concentration of indicators required for cell loading must be determined empirically.

Note The nonionic detergent Pluronic® F-127 is sometimes used to increase the aqueous solubility of Indo-1 AM. A variety of Pluronic® F-127 solutions can be purchased from AAT Bioquest.

Note If your cells contain organic anion-transporters, probenecid (1-2 mM) may be added to the dye working solution (final in well concentration will be 0.5-1 mM) to reduce leakage of the de-esterified indicators. A variety of ReadiUse™ probenecid products, including water-soluble, sodium salt, and stabilized solution, can be purchased from AAT Bioquest.

SAMPLE EXPERIMENTAL PROTOCOL

Following is our recommended protocol for loading AM esters into live cells. This protocol only provides a guideline and should be modified according to your specific needs.

1. Prepare cells in growth medium overnight.
2. On the next day, add 1X Indo-1 AM working solution into your cell plate.

Note If your compound(s) interfere with the serum, replace the growth medium with fresh HHBS buffer before dye-loading.

3. Incubate the dye-loaded plate in a cell incubator at 37 °C for 30 to 60 minutes.

Note Incubating the dye for longer than 1 hour can improve signal intensities in certain cell lines.

4. Replace the dye working solution with HHBS or buffer of your choice (containing an anion transporter inhibitor, such as 1 mM probenecid, if applicable) to remove any excess probes.
5. Add the stimulant as desired and simultaneously measure fluorescence using either a fluorescence microscope equipped with an Indo-1 filter set or a fluorescence plate reader containing a programmable liquid handling system such as a FlexStation, at Ex/Em₁ = 340/400 nm no cutoff and Ex/Em₂ = 340/475 cutoff 455 nm.


EXAMPLE DATA ANALYSIS AND FIGURES


Figure 1. Fluorescence emission spectra of Indo-1 in solutions containing 0 to 39 μM free Ca²⁺.

DISCLAIMER

AAT Bioquest provides high-quality reagents and materials for research use only. For proper handling of potentially hazardous chemicals, please consult the

Safety Data Sheet (SDS) provided for the product. Chemical analysis and/or reverse engineering of any kit or its components is strictly prohibited without written permission from AAT Bioquest. Please call 408-733-1055 or email info@aatbio.com if you have any questions.